

Product: Wobble-notTM Low Insertion Force Serological Pipet - XX mL
Individually wrapped (paper/plastic), 200/cs

Product Number: 4090-0XXX

Lot Number: XXXXXXXX

Manufacture Date: day-month-year

Date of Expiry: day-month-year

Pipet material: Virgin polystyrene.

Plug material: Polypropylene fibers.

Pipets are manufactured and sealed in sleeves inside a US 209 E Class 100 cleanroom (Equivalent to ISO 14644-1 ISO 5).

Sterilization: <10⁻⁶ SAL

Pre-sterilization is by gamma radiation with post-sterilization verification. This lot has been processed and tested in accordance with ISO 11137-2. Product samples are bio-burdened.

RNase/DNase: RNase <10⁻⁹ Kunitz units/ μ L, DNase < 10⁻⁷ Kunitz units/ μ L

This product has been tested and is free of any detectable RNase and DNase.

Pyrogen: <0.1 EU/mL

Endotoxin levels are tested by the limulus amoebocyte lysate (LAL) gel clot assay (sensitivity 0.1 EU/mL).

Cytotoxicity: Pass

ISO 10993-5.2009. Tested for reactivity in cell culture with ATCC[®] #CCL-1TM, NCTC clone 929.

Volumetric Accuracy: \pm 2%

This product has been tested and is accurate within \pm 2% at stated full volume.

Animal Content:

Resins, inks and packaging materials are not of animal origin and do not contain, make use of, or involve, at any point in the manufacturing process, raw materials of animal origin, including animal proteins.

Signed on behalf of the Quality Department:



Jeff Calhoun
V.P. Product Development